

CB TEST CERTIFICATE

Product	General Medical Devices USB-Isolator
Name and address of the applicant	<b>IFTOOLS GmbH</b> Matterhornstr. 17 65199 Wiesbaden GERMANY
Name and address of the manufacturer	IFTOOLS GmbH Matterhornstr. 17, 65199 Wiesbaden, GERMANY
Name and address of the factory	IFTOOLS GmbH Matterhornstr. 17, 65199 Wiesbaden, GERMANY
Ratings and principal characteristics	Rated input voltage: 5Vdc Rated input power: 900mA Protection class II Ingress protection: IP30
Model/type Ref.	ISOUSB-PLUS-CABLE ISOUSB-PLUS-BOX
Additional information (if necessary)	The following parts are excluded: IEC 60601-1:2005/AMD2:2020 clause 11.7 and 17;
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1:2005/AMD2:2020
as shown in the Test Report Ref. No. which forms part of this certificate	713266823



( Paul Bielawski )